

Specifications

Max. User Weight:	125kg (275 lbs)
Activity Level:	3
Size Range:	22 to 30
Component Weight:	2.6kg (5 lbs 11.7 oz) [†]
Build Height:	470mm to 565mm (18 ½" to 22 ¼")
Heel Height:	10mm

LINX | INTEGRATED LIMB SYSTEM



Order Example

Product Code	Size	Side	Spring Set
LINX	25	L	3

For dark tone add suffix D.
 Example: foot size 25 left, spring rating 3
 Proximal adaptor ordered separately

Adaptors

Rotating Pyramid with Shift	239017
Non-rotating Pyramid with Shift	239089
Rotating Female Pyramid with Shift	189128
M36 Threaded Adapter	239092
Long Pylon Kit, adds up to 4 ½"	339965



Suitable for outdoor use

Activity Level 3 has the ability or potential for ambulation with variable cadence. Typical of the community ambulator who has the ability to traverse most environmental barriers and may have vocational, therapeutic, or exercise activity that demands prosthetic utilisation beyond simple locomotion.

Activity

3

User Weight

44-52	53-59	60-68	69-77	78-88	89-100	101-116	117-125	kg
100-115	116-130	131-150	151-170	171-195	196-220	221-255	256-275	lbs

1	2	3	4	5	6	7	8
---	---	---	---	---	---	---	---

Spring set



[†]Component weight shown is for a size 26cm without foot shell.

References

- Hungerford DS, Cockin J. Fate of the retained lower limb joints in Second World War amputees. J Bone Joint Surg. 1975; 57(1):111.
- A.R. De Asha, R. Munjal, J. Kulkani, J.G. Buckley. Walking speed related joint kinetic alterations in trans-tibial amputees: impact of hydraulic 'ankle' damping. Journal of Neuro Engineering and Rehabilitation (JNER) 2013; 10:107
- Moore R. Effect on Stance Phase Timing Asymmetry in Individuals with Amputation Using Hydraulic Ankle Units. JPO: Journal of Prosthetics and Orthotics. 2016; 28(1):44-8.

Follow Endolite:

 @EndoliteUS | endolite.com
 /EndoliteUS | info@endolite.com

Endolite is a Blatchford Group Company. Endolite Corporate Headquarters, 1031 Byers Road, Miamisburg, Ohio 45342
t: 800.548.3534 f: 800.929.3636



An annual visual inspection is recommended. Check for visual defects that may affect proper function.
Maintenance must be carried out by competent personnel.

M00456US Iss4 11/17. Information correct at time of print.

